

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102119\
 Data File : VN058890.D
 Acq On : 21 Oct 2019 15:51
 Operator : JC/SP
 Sample : K5146-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-02-101819

Quant Time: Oct 22 04:38:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	402190	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	649142	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	605536	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	234764	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	293482	49.75	ug/l	0.00
Spiked Amount						
			Recovery	=	99.50%	
35) Dibromofluoromethane	7.57	113	209640	49.12	ug/l	0.00
Spiked Amount						
			Recovery	=	98.24%	
50) Toluene-d8	10.08	98	828263	49.99	ug/l	0.00
Spiked Amount						
			Recovery	=	99.98%	
62) 4-Bromofluorobenzene	12.40	95	277080	45.54	ug/l	0.00
Spiked Amount						
			Recovery	=	91.08%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.80	43	18051	6.193	ug/l	98
18) Methyl Acetate	4.31	43	9006	1.359	ug/l	96
20) Methylene Chloride	4.53	84	92375	18.560	ug/l	97
43) Isopropyl Acetate	8.15	43	106577	9.177	ug/l #	91
95) Naphthalene	15.13	128	29054	4.290	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102119\
Data File : VN058890.D
Acq On : 21 Oct 2019 15:51
Operator : JC/SP
Sample : K5146-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-02-101819

Quant Time: Oct 22 04:38:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

