

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102120\
 Data File : VN064220.D
 Acq On : 21 Oct 2020 14:42
 Operator : JC/MD
 Sample : L4408-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP-1

Quant Time: Oct 21 18:07:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 21 12:48:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	256140	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	448441	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	452443	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	201066	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	197075	51.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.62%	
35) Dibromofluoromethane	7.55	113	149082	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	
50) Toluene-d8	10.06	98	561387	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
62) 4-Bromofluorobenzene	12.38	95	250700	57.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.36%	
Target Compounds						
16) Acetone	3.79	43	10813	9.935	ug/l	96
17) Carbon Disulfide	4.01	76	11162	1.630	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102120\
Data File : VN064220.D
Acq On : 21 Oct 2020 14:42
Operator : JC/MD
Sample : L4408-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP-1

Quant Time: Oct 21 18:07:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 21 12:48:40 2020
Response via : Initial Calibration

