

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102120\
 Data File : VN064221.D
 Acq On : 21 Oct 2020 15:06
 Operator : JC/MD
 Sample : L4418-02DL 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB-2020-WW-000-02DL

Manual Integrations
 APPROVED

MMDadoda
 10/22/2020 11:28:36 AM

Quant Time: Oct 22 01:55:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 21 12:48:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	263133	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	454287	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	454859	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	207775	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	197864	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
35) Dibromofluoromethane	7.55	113	149443	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
50) Toluene-d8	10.06	98	565592	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
62) 4-Bromofluorobenzene	12.38	95	248101	55.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.72%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.79	96	5396m	1.704	ug/l	
44) Trichloroethene	8.80	130	10126	2.972	ug/l	96
64) Tetrachloroethene	10.60	164	428214	116.938	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102120\
Data File : VN064221.D
Acq On : 21 Oct 2020 15:06
Operator : JC/MD
Sample : L4418-02DL 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB-2020-WW-000-02DL

Manual Integrations
APPROVED

MMDadoda
10/22/2020 11:28:36 AM

Quant Time: Oct 22 01:55:06 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 21 12:48:40 2020
Response via : Initial Calibration

