

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102120\
 Data File : VN064228.D
 Acq On : 21 Oct 2020 17:56
 Operator : JC/MD
 Sample : L4413-01 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30688

Quant Time: Oct 22 02:24:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 21 12:48:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	273807	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	470496	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	472375	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	210713	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	208910	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
35) Dibromofluoromethane	7.55	113	158508	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
50) Toluene-d8	10.06	98	593168	52.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.54%	
62) 4-Bromofluorobenzene	12.38	95	257890	56.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102120\
Data File : VN064228.D
Acq On : 21 Oct 2020 17:56
Operator : JC/MD
Sample : L4413-01 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30688

Quant Time: Oct 22 02:24:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 21 12:48:40 2020
Response via : Initial Calibration

