

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102120\
 Data File : VN064231.D
 Acq On : 21 Oct 2020 19:07
 Operator : JC/MD
 Sample : L4434-05 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 16743

Quant Time: Oct 22 02:31:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 21 12:48:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	249369	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	438777	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	448019	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	202889	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	194350	52.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.96%	
35) Dibromofluoromethane	7.55	113	148250	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
50) Toluene-d8	10.06	98	554741	52.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.84%	
62) 4-Bromofluorobenzene	12.38	95	249835	58.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.48%	
Target Compounds						
16) Acetone	3.78	43	86022	81.183	ug/l	Qvalue 100
25) 2-Butanone	6.79	43	112110	66.594	ug/l	98
59) 2-Hexanone	10.73	43	7387	2.557	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102120\
Data File : VN064231.D
Acq On : 21 Oct 2020 19:07
Operator : JC/MD
Sample : L4434-05 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
16743

Quant Time: Oct 22 02:31:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 21 12:48:40 2020
Response via : Initial Calibration

