

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102120\
 Data File : VN064233.D
 Acq On : 21 Oct 2020 19:54
 Operator : JC/MD
 Sample : PB132513TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132513TB

Quant Time: Oct 22 02:36:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 21 12:48:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	258859	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	455029	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	464470	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	209681	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	201636	52.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.90%	
35) Dibromofluoromethane	7.55	113	151947	50.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.46%	
50) Toluene-d8	10.06	98	571622	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
62) 4-Bromofluorobenzene	12.38	95	258501	58.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.22%	
Target Compounds						
16) Acetone	3.79	43	13539	12.309	ug/l	100
18) Methyl Acetate	4.27	43	3753	1.339	ug/l #	65
20) Methylene Chloride	4.50	84	7330	2.499	ug/l	87
43) Isopropyl Acetate	8.13	43	93626	12.197	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102120\
Data File : VN064233.D
Acq On : 21 Oct 2020 19:54
Operator : JC/MD
Sample : PB132513TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132513TB

Quant Time: Oct 22 02:36:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 21 12:48:40 2020
Response via : Initial Calibration

