

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102120\
 Data File : VN064238.D
 Acq On : 22 Oct 2020 00:35
 Operator : JC/MD
 Sample : VN1021WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBL02

Quant Time: Oct 22 02:50:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 21 12:48:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	252048	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	444889	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	442837	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	202140	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.98	65	192878	51.53	ug/l	0.00
Spiked Amount						
			Recovery	=	103.06%	
35) Dibromofluoromethane	7.54	113	146889	49.66	ug/l	0.00
Spiked Amount						
			Recovery	=	99.32%	
50) Toluene-d8	10.06	98	551204	51.37	ug/l	0.00
Spiked Amount						
			Recovery	=	102.74%	
62) 4-Bromofluorobenzene	12.37	95	243553	55.99	ug/l	0.00
Spiked Amount						
			Recovery	=	111.98%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102120\
Data File : VN064238.D
Acq On : 22 Oct 2020 00:35
Operator : JC/MD
Sample : VN1021WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBL02

Quant Time: Oct 22 02:50:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 21 12:48:40 2020
Response via : Initial Calibration

