

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102120\
 Data File : VN064247.D
 Acq On : 22 Oct 2020 5:00
 Operator : JC/MD
 Sample : L4467-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-240

Quant Time: Oct 22 06:08:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 21 12:48:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	255039	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	449957	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	451907	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	210863	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	199347	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
35) Dibromofluoromethane	7.55	113	151614	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
50) Toluene-d8	10.06	98	569591	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
62) 4-Bromofluorobenzene	12.38	95	249719	56.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.52%	
Target Compounds						
16) Acetone	3.77	43	13949	12.872	ug/l	97
18) Methyl Acetate	4.28	43	3740	1.354	ug/l #	71
20) Methylene Chloride	4.50	84	5844	2.022	ug/l	89
43) Isopropyl Acetate	8.12	43	75062	9.889	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102120\
Data File : VN064247.D
Acq On : 22 Oct 2020 5:00
Operator : JC/MD
Sample : L4467-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-240

Quant Time: Oct 22 06:08:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 21 12:48:40 2020
Response via : Initial Calibration

