

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN102120\
 Data File : VN064237.D
 Acq On : 21 Oct 2020 23:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 22 02:45:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 21 12:48:40 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.547	0.587	-7.3	78	0.00
3 P	Chloromethane	0.578	0.644	-11.4	84	0.00
4 C	Vinyl Chloride	0.573	0.596	-4.0#	80	0.00
5 T	Bromomethane	0.373	0.418	-12.1	89	0.00
6 T	Chloroethane	0.349	0.371	-6.3	84	0.00
7 T	Trichlorofluoromethane	0.942	0.987	-4.8	80	0.00
8 T	Diethyl Ether	0.312	0.373	-19.6	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.487	0.476	2.3	76	0.00
10 T	Methyl Iodide	0.693	0.752	-8.5	85	0.00
11 T	Tert butyl alcohol	0.099	0.115	-16.2	93	0.00
12 CM	1,1-Dichloroethene	0.469	0.480	-2.3#	80	0.00
13 T	Acrolein	0.062	0.055	11.3	76	0.00
14 T	Allyl chloride	0.938	0.955	-1.8	79	0.00
15 T	Acrylonitrile	0.206	0.250	-21.4	95	0.00
16 T	Acetone	0.212	0.259	-22.2	97	0.00
17 T	Carbon Disulfide	1.337	1.365	-2.1	79	0.00
18 T	Methyl Acetate	0.541	0.646	-19.4	98	0.00
19 T	Methyl tert-butyl Ether	1.611	1.940	-20.4	94	0.00
20 T	Methylene Chloride	0.567	0.640	-12.9	93	0.00
21 T	trans-1,2-Dichloroethene	0.517	0.555	-7.4	85	0.00
22 T	Diisopropyl ether	1.744	2.002	-14.8	93	0.00
23 T	Vinyl Acetate	1.555	1.862	-19.7	93	0.00
24 P	1,1-Dichloroethane	1.010	1.156	-14.5	90	0.00
25 T	2-Butanone	0.338	0.392	-16.0	93	0.00
26 T	2,2-Dichloropropane	1.006	0.912	9.3	72	0.00
27 T	cis-1,2-Dichloroethene	0.602	0.677	-12.5	90	0.00
28 T	Bromochloromethane	0.480	0.563	-17.3	95	0.00
29 T	Tetrahydrofuran	0.202	0.242	-19.8	94	0.00
30 C	Chloroform	1.077	1.230	-14.2#	92	0.00
31 T	Cyclohexane	0.880	0.803	8.7	78	0.00
32 T	1,1,1-Trichloroethane	1.009	1.092	-8.2	84	0.00
33 S	1,2-Dichloroethane-d4	0.743	0.833	-12.1	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.332	0.333	-0.3	90	0.00
36 T	1,1-Dichloropropene	0.498	0.483	3.0	83	0.00
37 T	Ethyl Acetate	0.453	0.494	-9.1	92	0.00
38 T	Carbon Tetrachloride	0.567	0.544	4.1	80	0.00
39 T	Methylcyclohexane	0.453	0.419	7.5	76	0.00
40 TM	Benzene	1.376	1.418	-3.1	88	0.00
41 T	Methacrylonitrile	0.237	0.205	13.5	83	0.00
42 TM	1,2-Dichloroethane	0.589	0.649	-10.2	95	0.00
43 T	Isopropyl Acetate	0.843	0.951	-12.8	97	0.00
44 TM	Trichloroethene	0.375	0.375	0.0	85	0.00
45 C	1,2-Dichloropropane	0.367	0.396	-7.9#	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN102120\
 Data File : VN064237.D
 Acq On : 21 Oct 2020 23:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 22 02:45:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 21 12:48:40 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.247	0.278	-12.6	91	0.00
47 T	Bromodichloromethane	0.545	0.618	-13.4	93	0.00
48 T	Methyl methacrylate	0.441	0.474	-7.5	93	0.00
49 T	1,4-Dioxane	0.005	0.005	0.0	93	0.00
50 S	Toluene-d8	1.206	1.161	3.7	83	0.00
51 T	4-Methyl-2-Pentanone	0.451	0.503	-11.5	93	0.00
52 CM	Toluene	0.880	0.897	-1.9#	86	0.00
53 T	t-1,3-Dichloropropene	0.605	0.662	-9.4	90	0.00
54 T	cis-1,3-Dichloropropene	0.637	0.680	-6.8	90	0.00
55 T	1,1,2-Trichloroethane	0.333	0.377	-13.2	94	0.00
56 T	Ethyl methacrylate	0.514	0.584	-13.6	93	0.00
57 T	1,3-Dichloropropane	0.578	0.648	-12.1	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.265	0.296	-11.7	92	0.00
59 T	2-Hexanone	0.329	0.366	-11.2	91	0.00
60 T	Dibromochloromethane	0.430	0.470	-9.3	93	0.00
61 T	1,2-Dibromoethane	0.351	0.402	-14.5	94	0.00
62 S	4-Bromofluorobenzene	0.489	0.469	4.1	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.403	0.399	1.0	85	0.00
65 PM	Chlorobenzene	1.035	1.088	-5.1	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.411	0.441	-7.3	90	0.00
67 C	Ethyl Benzene	1.902	1.931	-1.5#	84	0.00
68 T	m/p-Xylenes	0.705	0.711	-0.9	85	0.00
69 T	o-Xylene	0.682	0.698	-2.3	84	0.00
70 T	Styrene	1.142	1.209	-5.9	86	0.00
71 P	Bromoform	0.317	0.353	-11.4	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.813	3.733	2.1	81	0.00
74 T	N-amyl acetate	1.648	1.858	-12.7	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.012	1.125	-11.2	93	0.00
76 T	1,2,3-Trichloropropane	1.054	1.073	-1.8	91	0.00
77 T	Bromobenzene	0.900	0.965	-7.2	89	0.00
78 T	n-propylbenzene	4.300	4.253	1.1	81	0.00
79 T	2-Chlorotoluene	2.641	2.696	-2.1	84	0.00
80 T	1,3,5-Trimethylbenzene	3.224	3.225	-0.0	82	0.00
81 T	trans-1,4-Dichloro-2-butene	0.418	0.437	-4.5	87	0.00
82 T	4-Chlorotoluene	2.755	2.859	-3.8	84	0.00
83 T	tert-Butylbenzene	2.643	2.563	3.0	80	0.00
84 T	1,2,4-Trimethylbenzene	3.209	3.281	-2.2	83	0.00
85 T	sec-Butylbenzene	3.332	3.273	1.8	80	0.00
86 T	p-Isopropyltoluene	3.035	3.052	-0.6	81	0.00
87 T	1,3-Dichlorobenzene	1.602	1.657	-3.4	85	0.00
88 T	1,4-Dichlorobenzene	1.640	1.677	-2.3	85	0.00
89 T	n-Butylbenzene	2.720	2.657	2.3	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN102120\
Data File : VN064237.D
Acq On : 21 Oct 2020 23:56
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 22 02:45:28 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 21 12:48:40 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.550	0.520	5.5	79	0.00
91 T	1,2-Dichlorobenzene	1.572	1.644	-4.6	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.252	0.274	-8.7	89	0.00
93 T	1,2,4-Trichlorobenzene	0.881	0.953	-8.2	84	0.00
94 T	Hexachlorobutadiene	0.475	0.412	13.3	73	0.00
95 T	Naphthalene	2.640	3.019	-14.4	90	0.00
96 T	1,2,3-Trichlorobenzene	0.873	0.913	-4.6	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6