

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
 Data File : VN069083.D
 Acq On : 21 Oct 2021 13:06
 Operator : JC/MD
 Sample : M4271-01
 Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-01-101821

Quant Time: Oct 21 13:32:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	315896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	558841	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	495395	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	210715	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	210433	50.524	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.040%
35) Dibromofluoromethane	8.019	113	161240	49.591	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.180%
50) Toluene-d8	10.441	98	689110	52.970	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.940%
62) 4-Bromofluorobenzene	12.734	95	243591	49.629	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
Data File : VN069083.D
Acq On : 21 Oct 2021 13:06
Operator : JC/MD
Sample : M4271-01
Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-01-101821

Quant Time: Oct 21 13:32:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

