

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
 Data File : VN069085.D
 Acq On : 21 Oct 2021 13:56
 Operator : JC/MD
 Sample : M4253-04 200X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 129

Quant Time: Oct 22 01:32:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	347161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	612986	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	576726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	210564	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	229960	50.240	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.480%			
35) Dibromofluoromethane	8.024	113	179362	50.292	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.580%			
50) Toluene-d8	10.446	98	758355	53.144	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.280%			
62) 4-Bromofluorobenzene	12.734	95	268183	49.813	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 99.620%			
Target Compounds					Qvalue	
16) Acetone	4.291	43	143569	84.051	ug/l	98
20) Methylene Chloride	5.103	84	2945	0.749	ug/l	84
40) Benzene	8.461	78	5292	0.332	ug/l	97
67) Ethyl Benzene	11.849	91	22244	1.097	ug/l	97
68) m/p-Xylenes	11.953	106	37385	4.865	ug/l	100
69) o-Xylene	12.283	106	18307	2.380	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
Data File : VN069085.D
Acq On : 21 Oct 2021 13:56
Operator : JC/MD
Sample : M4253-04 200X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
129

Quant Time: Oct 22 01:32:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

