

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
 Data File : VN069087.D
 Acq On : 21 Oct 2021 14:47
 Operator : JC/MD
 Sample : M4284-11 200X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30854

Quant Time: Oct 22 01:33:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	318125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	557397	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	522895	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	199444	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	209878	50.038	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 100.080%		
35) Dibromofluoromethane	8.027	113	164234	50.643	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 101.280%		
50) Toluene-d8	10.443	98	694540	53.526	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 107.060%		
62) 4-Bromofluorobenzene	12.733	95	250096	51.086	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 102.180%		
Target Compounds						Qvalue
16) Acetone	4.309	43	7190	4.593	ug/l #	85
30) Chloroform	7.820	83	19548	2.979	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
Data File : VN069087.D
Acq On : 21 Oct 2021 14:47
Operator : JC/MD
Sample : M4284-11 200X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30854

Quant Time: Oct 22 01:33:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

