

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
 Data File : VN069088.D
 Acq On : 21 Oct 2021 15:12
 Operator : JC/MD
 Sample : M4284-06 200X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20878-WATER

Quant Time: Oct 22 01:33:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	335097	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	587693	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	536256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	192338	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	219980	49.790	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.580%
35) Dibromofluoromethane	8.027	113	169962	49.707	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.420%
50) Toluene-d8	10.446	98	727748	53.194	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.380%
62) 4-Bromofluorobenzene	12.733	95	243292	47.134	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
Data File : VN069088.D
Acq On : 21 Oct 2021 15:12
Operator : JC/MD
Sample : M4284-06 200X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20878-WATER

Quant Time: Oct 22 01:33:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

