

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN102121\  
 Data File : VN069091.D  
 Acq On : 21 Oct 2021 16:28  
 Operator : JC/MD  
 Sample : M4284-16 200X  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 30849

Quant Time: Oct 22 01:33:53 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N101221W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 13 05:44:40 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)      |
|-----------------------------|--------|-------|----------|----------|-------|---------------|
| -----                       |        |       |          |          |       |               |
| Internal Standards          |        |       |          |          |       |               |
| 1) Pentafluorobenzene       | 8.088  | 168   | 317075   | 50.000   | ug/l  | 0.00          |
| 34) 1,4-Difluorobenzene     | 8.971  | 114   | 565479   | 50.000   | ug/l  | 0.00          |
| 63) Chlorobenzene-d5        | 11.749 | 117   | 531180   | 50.000   | ug/l  | 0.00          |
| 72) 1,4-Dichlorobenzene-d4  | 13.678 | 152   | 199049   | 50.000   | ug/l  | 0.00          |
| System Monitoring Compounds |        |       |          |          |       |               |
| 33) 1,2-Dichloroethane-d4   | 8.440  | 65    | 210295   | 50.303   | ug/l  | 0.00          |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =     | 100.600%      |
| 35) Dibromofluoromethane    | 8.027  | 113   | 162942   | 49.526   | ug/l  | 0.00          |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =     | 99.060%       |
| 50) Toluene-d8              | 10.446 | 98    | 704222   | 53.497   | ug/l  | 0.00          |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =     | 107.000%      |
| 62) 4-Bromofluorobenzene    | 12.734 | 95    | 246797   | 49.692   | ug/l  | 0.00          |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =     | 99.380%       |
| Target Compounds            |        |       |          |          |       |               |
| 16) Acetone                 | 4.293  | 43    | 34153    | 21.892   | ug/l  | Qvalue<br>100 |
| -----                       |        |       |          |          |       |               |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN102121\  
Data File : VN069091.D  
Acq On : 21 Oct 2021 16:28  
Operator : JC/MD  
Sample : M4284-16 200X  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
30849

Quant Time: Oct 22 01:33:53 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N101221W.M  
Quant Title : SW846 8260  
QLast Update : Wed Oct 13 05:44:40 2021  
Response via : Initial Calibration

