

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
 Data File : VN069099.D
 Acq On : 21 Oct 2021 19:49
 Operator : JC/MD
 Sample : PB140087TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB140087TB

Quant Time: Oct 22 01:35:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	327935	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	577780	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	568587	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	209870	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	216436	50.058	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.120%			
35) Dibromofluoromethane	8.024	113	168925	50.252	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.500%			
50) Toluene-d8	10.443	98	731366	54.376	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 108.760%			
62) 4-Bromofluorobenzene	12.733	95	261775	51.585	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 103.180%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	16059	9.953	ug/l	98
20) Methylene Chloride	5.103	84	15118	4.071	ug/l	96
43) Isopropyl Acetate	8.563	43	34745	3.863	ug/l #	92
95) Naphthalene	15.509	128	13034	5.048	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
Data File : VN069099.D
Acq On : 21 Oct 2021 19:49
Operator : JC/MD
Sample : PB140087TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB140087TB

Quant Time: Oct 22 01:35:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

