

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
 Data File : VN069109.D
 Acq On : 22 Oct 2021 00:51
 Operator : JC/MD
 Sample : M4307-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Oct 22 01:37:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	364643	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	625537	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	596119	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	225519	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	231383	48.128	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.260%
35) Dibromofluoromethane	8.027	113	183669	50.466	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.940%
50) Toluene-d8	10.443	98	773134	53.093	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.180%
62) 4-Bromofluorobenzene	12.733	95	276149	50.263	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.520%
Target Compounds						
					Qvalue	
6) Chloroethane	3.030	64	4715	1.767	ug/l	92
16) Acetone	4.304	43	6633	3.697	ug/l	98
95) Naphthalene	15.512	128	5487	4.296	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
Data File : VN069109.D
Acq On : 22 Oct 2021 00:51
Operator : JC/MD
Sample : M4307-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Oct 22 01:37:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

