

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
 Data File : VN069111.D
 Acq On : 22 Oct 2021 1:41
 Operator : JC/MD
 Sample : M4274-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2-BS

Quant Time: Oct 22 03:17:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	353035	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	604500	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	612432	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	224034	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	222147	47.726	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.460%		
35) Dibromofluoromethane	8.027	113	176422	50.162	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	100.320%		
50) Toluene-d8	10.446	98	769595	54.689	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	109.380%		
62) 4-Bromofluorobenzene	12.733	95	274316	51.667	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	103.340%		
Target Compounds					Qvalue	
16) Acetone	4.299	43	30784	17.722	ug/l	98
20) Methylene Chloride	5.109	84	26371	6.596	ug/l	96
43) Isopropyl Acetate	8.566	43	35964	3.822	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
Data File : VN069111.D
Acq On : 22 Oct 2021 1:41
Operator : JC/MD
Sample : M4274-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2-BS

Quant Time: Oct 22 03:17:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

