

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
 Data File : VN069113.D
 Acq On : 22 Oct 2021 2:31
 Operator : JC/MD
 Sample : M4274-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2-DS

Quant Time: Oct 22 03:17:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	349011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	606990	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	612247	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	229937	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	224604	48.810	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.620%
35) Dibromofluoromethane	8.024	113	176740	50.046	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.100%
50) Toluene-d8	10.443	98	764602	54.111	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.220%
62) 4-Bromofluorobenzene	12.734	95	278528	52.245	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.500%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	25648	14.936	ug/l	96
20) Methylene Chloride	5.103	84	27927	7.066	ug/l	90
43) Isopropyl Acetate	8.563	43	46077	4.877	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
Data File : VN069113.D
Acq On : 22 Oct 2021 2:31
Operator : JC/MD
Sample : M4274-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2-DS

Quant Time: Oct 22 03:17:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

