

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
 Data File : VN069117.D
 Acq On : 22 Oct 2021 4:12
 Operator : JC/MD
 Sample : M4285-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-GA-(1)

Quant Time: Oct 22 07:21:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	337490	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	593697	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	594595	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	221619	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	218737	49.158	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.320%
35) Dibromofluoromethane	8.024	113	172721	50.003	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.000%
50) Toluene-d8	10.446	98	744446	53.864	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.720%
62) 4-Bromofluorobenzene	12.733	95	271358	52.040	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.080%
Target Compounds						
					Qvalue	
16) Acetone	4.301	43	11232	6.764	ug/l	99
18) Methyl Acetate	4.870	43	10114	1.606	ug/l #	66
20) Methylene Chloride	5.101	84	12117	3.170	ug/l	96
43) Isopropyl Acetate	8.568	43	57805	6.255	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
Data File : VN069117.D
Acq On : 22 Oct 2021 4:12
Operator : JC/MD
Sample : M4285-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-GA-(1)

Quant Time: Oct 22 07:21:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

