

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
 Data File : VN069079.D
 Acq On : 21 Oct 2021 11:08
 Operator : JC/MD
 Sample : VN1021MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021MBL01

Quant Time: Oct 21 13:00:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	353048	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	615650	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	576531	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	207648	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	233053	50.067	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.140%
35) Dibromofluoromethane	8.024	113	182908	51.064	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.120%
50) Toluene-d8	10.443	98	767653	53.563	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.120%
62) 4-Bromofluorobenzene	12.734	95	264724	48.958	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.920%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102121\
Data File : VN069079.D
Acq On : 21 Oct 2021 11:08
Operator : JC/MD
Sample : VN1021MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1021MBL01

Quant Time: Oct 21 13:00:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.088 min Scan# 2274
Delta R.T. 0.002 min
Lab File: VN069079.D
Acq: 21 Oct 2021 11:08




Instrument :
MSVOA_N
ClientSampleId :
VN1021MBL01


Tgt Ion:168 Resp: 353048
Ion Ratio Lower Upper
168 100
99 62.1 48.9 73.3



#33
1,2-Dichloroethane-d4
Concen: 50.067 ug/l
RT: 8.437 min Scan# 2404
Delta R.T. -0.003 min
Lab File: VN069079.D
Acq: 21 Oct 2021 11:08

Tgt Ion: 65 Resp: 233053
Ion Ratio Lower Upper
65 100
67 51.5 0.0 103.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.968 min Scan# 2602
Delta R.T. 0.000 min
Lab File: VN069079.D
Acq: 21 Oct 2021 11:08




Instrument :
MSVOA_N
ClientSampleId :
VN1021MBL01


Tgt Ion:114 Resp: 615650
Ion Ratio Lower Upper
114 100
63 20.9 0.0 42.6
88 16.2 0.0 32.0



#35
Dibromofluoromethane
Concen: 51.064 ug/l
RT: 8.024 min Scan# 2250
Delta R.T. 0.003 min
Lab File: VN069079.D
Acq: 21 Oct 2021 11:08

Tgt Ion:113 Resp: 182908
Ion Ratio Lower Upper
113 100
111 102.6 83.2 124.8
192 17.8 15.0 22.4









#50
Toluene-d8
Concen: 53.563 ug/l
RT: 10.443 min Scan# 3152
Delta R.T. 0.000 min
Lab File: VN069079.D
Acq: 21 Oct 2021 11:08




Instrument :
MSVOA_N
ClientSampleId :
VN1021MBL01


Tgt Ion: 98 Resp: 767653
Ion Ratio Lower Upper
98 100
100 64.3 51.5 77.3



#62
4-Bromofluorobenzene
Concen: 48.958 ug/l
RT: 12.734 min Scan# 4006
Delta R.T. 0.000 min
Lab File: VN069079.D
Acq: 21 Oct 2021 11:08

Tgt Ion: 95 Resp: 264724
Ion Ratio Lower Upper
95 100
174 72.6 0.0 144.8
176 69.7 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.747 min Scan# 3638
Delta R.T. 0.000 min
Lab File: VN069079.D
Acq: 21 Oct 2021 11:08

Instrument :
MSVOA_N
ClientSampleId :
VN1021MBL01

Tgt Ion:117 Resp: 576531
Ion Ratio Lower Upper
117 100
82 57.5 46.4 69.6
119 32.2 26.0 39.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.678 min Scan# 4358
Delta R.T. 0.003 min
Lab File: VN069079.D
Acq: 21 Oct 2021 11:08

Tgt Ion:152 Resp: 207648
Ion Ratio Lower Upper
152 100
115 63.4 31.4 94.3
150 157.7 0.0 351.0

