

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
 Data File : VN074954.D
 Acq On : 21 Oct 2022 11:30
 Operator : JC\MD
 Sample : VN1021WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBL01

Quant Time: Oct 22 02:51:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 21 09:37:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	640755	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	1071206	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	1103094	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	610340	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	739491	43.437	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.880%
35) Dibromofluoromethane	8.171	113	556488	43.646	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.300%
50) Toluene-d8	10.571	98	2322975	47.486	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.980%
62) 4-Bromofluorobenzene	12.847	95	765207	44.345	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
Data File : VN074954.D
Acq On : 21 Oct 2022 11:30
Operator : JC\MD
Sample : VN1021WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBL01

Quant Time: Oct 22 02:51:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 21 09:37:47 2022
Response via : Initial Calibration

