

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
 Data File : VN074962.D
 Acq On : 21 Oct 2022 15:23
 Operator : JC\MD
 Sample : N5222-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Quant Time: Oct 22 02:53:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 21 09:37:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	800874	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	1322077	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	1301280	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	701006	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	860292	40.430	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	80.860%
35) Dibromofluoromethane	8.171	113	675975	42.957	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.920%
50) Toluene-d8	10.571	98	2738189	45.353	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.700%
62) 4-Bromofluorobenzene	12.847	95	939712	44.124	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.240%
Target Compounds						
					Qvalue	
30) Chloroform	7.977	83	21109	0.678	ug/l	88
52) Toluene	10.635	92	11205	0.281	ug/l	97
64) Tetrachloroethene	11.112	164	6016	0.565	ug/l #	89
67) Ethyl Benzene	11.965	91	80564	1.153	ug/l	97
68) m/p-Xylenes	12.076	106	79173	2.907	ug/l	98
78) n-propylbenzene	13.035	91	16115	0.240	ug/l	96
80) 1,3,5-Trimethylbenzene	13.176	105	18137	0.367	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	132186	2.646	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
Data File : VN074962.D
Acq On : 21 Oct 2022 15:23
Operator : JC\MD
Sample : N5222-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Quant Time: Oct 22 02:53:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 21 09:37:47 2022
Response via : Initial Calibration

