

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
 Data File : VN074964.D
 Acq On : 21 Oct 2022 16:11
 Operator : JC\MD
 Sample : N5222-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-7

Quant Time: Oct 22 02:54:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 21 09:37:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	716989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	1232131	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	1251390	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	667882	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	795894	41.780	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.560%
35) Dibromofluoromethane	8.177	113	620779	42.329	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.660%
50) Toluene-d8	10.571	98	2566304	45.609	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.220%
62) 4-Bromofluorobenzene	12.847	95	875818	44.126	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.260%
Target Compounds						
					Qvalue	
30) Chloroform	7.977	83	7389	0.265	ug/l #	71
64) Tetrachloroethene	11.094	164	4176	0.408	ug/l #	48
68) m/p-Xylenes	12.076	106	12978	0.495	ug/l	86
84) 1,2,4-Trimethylbenzene	13.488	105	19567	0.411	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
Data File : VN074964.D
Acq On : 21 Oct 2022 16:11
Operator : JC\MD
Sample : N5222-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-7

Quant Time: Oct 22 02:54:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 21 09:37:47 2022
Response via : Initial Calibration

