

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
 Data File : VN074965.D
 Acq On : 21 Oct 2022 16:36
 Operator : JC\MD
 Sample : N5222-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-8

Quant Time: Oct 22 02:54:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 21 09:37:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	670905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	1142692	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	1175432	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	623718	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	760500	42.664	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.320%
35) Dibromofluoromethane	8.177	113	580623	42.690	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.380%
50) Toluene-d8	10.571	98	2454183	47.030	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.060%
62) 4-Bromofluorobenzene	12.853	95	827870	44.975	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.960%
Target Compounds						
					Qvalue	
30) Chloroform	7.971	83	21464	0.822	ug/l #	79
68) m/p-Xylenes	12.082	106	9068	0.369	ug/l	82
84) 1,2,4-Trimethylbenzene	13.476	105	15291	0.344	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
Data File : VN074965.D
Acq On : 21 Oct 2022 16:36
Operator : JC\MD
Sample : N5222-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-8

Quant Time: Oct 22 02:54:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 21 09:37:47 2022
Response via : Initial Calibration

