

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
 Data File : VN074966.D
 Acq On : 21 Oct 2022 17:00
 Operator : JC\MD
 Sample : N5222-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-5

Quant Time: Oct 22 02:54:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 21 09:37:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	665965	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	1120384	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	1131212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	616660	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	737314	41.670	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.340%
35) Dibromofluoromethane	8.171	113	564853	42.357	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.720%
50) Toluene-d8	10.571	98	2378604	46.489	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.980%
62) 4-Bromofluorobenzene	12.853	95	811499	44.964	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.920%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	5.818	73	192383	4.478	ug/l #	75
40) Benzene	8.612	78	20315	0.392	ug/l #	81
64) Tetrachloroethene	11.106	164	12507	1.352	ug/l #	75
67) Ethyl Benzene	11.965	91	15129	0.249	ug/l #	83
84) 1,2,4-Trimethylbenzene	13.488	105	10864	0.247	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
Data File : VN074966.D
Acq On : 21 Oct 2022 17:00
Operator : JC\MD
Sample : N5222-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-5

Quant Time: Oct 22 02:54:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 21 09:37:47 2022
Response via : Initial Calibration

