

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
 Data File : VN074967.D
 Acq On : 21 Oct 2022 17:24
 Operator : JC\MD
 Sample : N5222-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-10D

Quant Time: Oct 22 02:55:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 21 09:37:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	665780	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	1117553	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	1139536	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	606021	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	733374	41.459	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.920%
35) Dibromofluoromethane	8.171	113	580726	43.658	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.320%
50) Toluene-d8	10.571	98	2356056	46.165	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.320%
62) 4-Bromofluorobenzene	12.847	95	809047	44.941	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.880%
Target Compounds						
					Qvalue	
30) Chloroform	7.971	83	207630	8.016	ug/l	95
64) Tetrachloroethene	11.106	164	60527	6.495	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
Data File : VN074967.D
Acq On : 21 Oct 2022 17:24
Operator : JC\MD
Sample : N5222-13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-10D

Quant Time: Oct 22 02:55:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 21 09:37:47 2022
Response via : Initial Calibration

