

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
 Data File : VN074972.D
 Acq On : 21 Oct 2022 19:25
 Operator : JC\MD
 Sample : N5222-18
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-16D

Quant Time: Oct 22 02:56:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 21 09:37:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	744415	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	1245769	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	1259115	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	687204	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	832235	42.078	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.160%
35) Dibromofluoromethane	8.177	113	646765	43.618	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.240%
50) Toluene-d8	10.570	98	2650973	46.598	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.200%
62) 4-Bromofluorobenzene	12.853	95	923258	46.007	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.020%
Target Compounds						
					Qvalue	
30) Chloroform	7.971	83	126330	4.362	ug/l	95
64) Tetrachloroethene	11.100	164	17316	1.682	ug/l #	78
84) 1,2,4-Trimethylbenzene	13.488	105	33338	0.681	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
Data File : VN074972.D
Acq On : 21 Oct 2022 19:25
Operator : JC\MD
Sample : N5222-18
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-16D

Quant Time: Oct 22 02:56:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 21 09:37:47 2022
Response via : Initial Calibration

