

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
 Data File : VN074974.D
 Acq On : 21 Oct 2022 20:13
 Operator : JC\MD
 Sample : N5222-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Oct 22 02:56:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 21 09:37:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	690736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	1192882	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	1295714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	791629	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	806192	43.929	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.860%
35) Dibromofluoromethane	8.177	113	604788	42.596	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.200%
50) Toluene-d8	10.571	98	2624913	48.185	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.380%
62) 4-Bromofluorobenzene	12.853	95	967411	50.345	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
Data File : VN074974.D
Acq On : 21 Oct 2022 20:13
Operator : JC\MD
Sample : N5222-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Oct 22 02:56:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 21 09:37:47 2022
Response via : Initial Calibration

