



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102122\
 Data File : VN074951.D
 Acq On : 21 Oct 2022 08:52
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
 Title : SW846 8260
 Last Update : Fri Oct 21 09:37:47 2022



AutoFind: Scans 1834, 1835, 1836; Background Corrected with Scan 1827

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	18696	PASS
75	95	30	60	57.6	52141	PASS
95	95	100	100	100.0	90520	PASS
96	95	5	9	6.1	5541	PASS
173	174	0.00	2	1.1	753	PASS
174	95	50	100	73.3	66352	PASS
175	174	5	9	7.1	4715	PASS
176	174	95	101	99.5	65997	PASS
177	176	5	9	6.6	4336	PASS

Average of 12.841 to 12.853 min.: VN074951.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.90	1081	47.05	1153	57.80	85	69.90	45
36.30	149	47.75	133	59.85	952	70.05	81
36.95	4604	49.05	4838	61.00	6397	71.70	22
38.05	4148	50.00	18696	62.05	4707	72.10	24
38.95	1691	51.05	6821	62.95	3597	73.00	5012
41.00	303	51.85	430	63.90	817	74.05	20419
43.95	1451	54.80	74	65.10	893	75.00	52141
44.90	310	55.05	216	66.20	133	75.95	3929
45.10	733	56.00	771	66.95	272	76.95	242
46.10	60	56.15	1002	68.05	11259	78.00	437
46.70	276	57.00	3528	69.00	9830	78.95	4652

Average of 12.841 to 12.853 min.: VN074951.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
79.95	1497	92.05	2756	115.80	134	136.90	127
80.95	5021	92.90	603	116.10	414	137.20	142
81.85	1036	93.10	3228	116.95	940	138.75	135
83.20	52	94.00	12715	117.95	290	139.80	80
86.90	2662	95.00	90520	118.85	663	140.85	534
87.10	1823	96.05	5541	125.90	102	141.10	392
87.90	873	103.85	488	127.85	573	141.85	316
88.05	3956	104.10	202	128.95	266	142.95	1399
90.10	50	104.85	266	129.60	135	146.00	226
90.70	108	105.80	550	129.90	209	146.75	251
91.05	505	114.60	88	134.85	177	147.90	311

Average of 12.841 to 12.853 min.: VN074951.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
148.80	51	172.85	753				
149.80	154	174.00	66352				
153.90	166	174.95	4715				
155.05	288	176.00	65997				
156.75	132	176.95	4336				
158.80	147	177.80	113				
160.80	96	207.00	136				
162.60	52	280.90	118				
171.50	97	281.15	290				
172.05	710	282.30	144				
172.50	324	283.20	52				