

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084437.D
 Acq On : 21 Oct 2024 15:41
 Operator : JC\MD
 Sample : P4440-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BPOW6-11-TB-20241016

Quant Time: Oct 22 07:54:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.224	168	177910	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	306096	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	264394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103232	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	125403	47.540	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.080%		
35) Dibromofluoromethane	8.165	113	102688	50.887	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.780%		
50) Toluene-d8	10.565	98	364855	49.143	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.280%		
62) 4-Bromofluorobenzene	12.847	95	123415	45.629	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	91.260%		
Target Compounds					Qvalue	
16) Acetone	4.453	43	1153	1.018	ug/l #	62
20) Methylene Chloride	5.277	84	983	0.427	ug/l #	86
31) Cyclohexane	8.218	56	4150	1.062	ug/l #	1
36) 1,1-Dichloropropene	8.224	75	15970	5.448	ug/l #	50
38) Carbon Tetrachloride	8.224	117	19559	6.083	ug/l #	17
51) 4-Methyl-2-Pentanone	10.565	43	1520	0.531	ug/l #	1
71) Bromoform	12.841	173	773	0.520	ug/l #	1
76) 1,2,3-Trichloropropane	12.847	75	64419	30.566	ug/l #	1
81) trans-1,4-Dichloro-2-b...	12.847	75	64419	73.282	ug/l #	11


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084437.D
Acq On : 21 Oct 2024 15:41
Operator : JC\MD
Sample : P4440-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016

Quant Time: Oct 22 07:54:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41




Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016


Tgt Ion:168 Resp: 177910
Ion Ratio Lower Upper
168 100
99 61.1 54.2 81.2



#16
Acetone
Concen: 1.018 ug/l
RT: 4.453 min Scan# 425
Delta R.T. 0.029 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41

Tgt Ion: 43 Resp: 1153
Ion Ratio Lower Upper
43 100
58 51.3 24.4 36.6#









#20
Methylene Chloride
Concen: 0.427 ug/l
RT: 5.277 min Scan# 51
Delta R.T. -0.000 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41




Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016


Tgt Ion:	84	Resp:	983
Ion	Ratio	Lower	Upper
84	100		
49	150.7	107.2	160.8
51	50.2	31.8	47.8#
86	57.6	52.9	79.3



#31
Cyclohexane
Concen: 1.062 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.035 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41

Tgt Ion:	56	Resp:	4150
Ion	Ratio	Lower	Upper
56	100		
69	158.8	23.8	35.6#
84	131.9	68.8	103.2#





#33

1,2-Dichloroethane-d4

Concen: 47.540 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VN084437.D

Acq: 21 Oct 2024 15:41

Instrument :

MSVOA_N

ClientSampleId :





BPOW6-11-TB-20241016

Tgt Ion: 65 Resp: 125403

Ion Ratio Lower Upper

65 100

67 50.4 0.0 102.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN084437.D

Acq: 21 Oct 2024 15:41




Tgt Ion: 114 Resp: 306096

Ion Ratio Lower Upper

114 100

63 20.8 0.0 43.8

88 16.2 0.0 31.6





#35

Dibromofluoromethane

Concen: 50.887 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN084437.D

Acq: 21 Oct 2024 15:41

Instrument :

MSVOA_N

ClientSampleId :

BPOW6-11-TB-20241016





Tgt Ion:113 Resp: 102688

Ion Ratio Lower Upper

113 100

111 101.7 83.3 124.9

192 17.6 13.5 20.3



#36

1,1-Dichloropropene

Concen: 5.448 ug/l

RT: 8.224 min Scan# 1066

Delta R.T. -0.147 min

Lab File: VN084437.D

Acq: 21 Oct 2024 15:41




Tgt Ion: 75 Resp: 15970


Ion Ratio Lower Upper

75 100

110 8.1 16.9 50.7#

77 0.0 24.7 37.1#









#38
Carbon Tetrachloride
Concen: 6.083 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.135 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41




Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016


Tgt Ion:117 Resp: 19559
Ion Ratio Lower Upper
117 100
119 7.3 78.7 118.1#
121 0.0 24.4 36.6#



#50
Toluene-d8
Concen: 49.143 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41

Tgt Ion: 98 Resp: 364855
Ion Ratio Lower Upper
98 100
100 64.5 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 0.531 ug/l
RT: 10.565 min Scan# 1443
Delta R.T. 0.123 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41




Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016


Tgt Ion: 43 Resp: 1520
Ion Ratio Lower Upper
43 100
58 183.5 31.6 47.4#



#62
4-Bromofluorobenzene
Concen: 45.629 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41

Tgt Ion: 95 Resp: 123415
Ion Ratio Lower Upper
95 100
174 74.5 0.0 145.2
176 70.5 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN084437.D

Acq: 21 Oct 2024 15:41

Instrument :

MSVOA_N

ClientSampleId :

BPOW6-11-TB-20241016





Tgt Ion:117 Resp: 264394

Ion Ratio Lower Upper

117 100

82 56.5 47.2 70.8

119 31.3 25.4 38.0



#71

Bromoform

Concen: 0.520 ug/l

RT: 12.841 min Scan# 1851

Delta R.T. 0.265 min

Lab File: VN084437.D

Acq: 21 Oct 2024 15:41




Tgt Ion:173 Resp: 773


Ion Ratio Lower Upper

173 100

175 682.0 23.5 70.6#

254 0.0 0.0 0.0









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2012
Delta R.T. 0.006 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41




Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016


Tgt Ion:152 Resp: 103232
Ion Ratio Lower Upper
152 100
115 63.2 31.3 93.9
150 159.8 0.0 349.8



#76
1,2,3-Trichloropropane
Concen: 30.566 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.141 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41


Tgt Ion: 75 Resp: 64419
Ion Ratio Lower Upper
75 100
77 1.1 96.9 290.7#





#81
trans-1,4-Dichloro-2-butene
Concen: 73.282 ug/l
RT: 12.847 min Scan# 11
Delta R.T. 0.112 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016



Tgt Ion: 75 Resp: 64419
Ion Ratio Lower Upper
75 100
53 0.0 78.0 117.0#
89 0.0 35.5 53.3#

