

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084435.D
 Acq On : 21 Oct 2024 14:53
 Operator : JC\MD
 Sample : P4438-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VPB190-HYD-20241016

Quant Time: Oct 22 01:34:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	171057	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	299485	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	256223	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	104212	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	121285	47.821	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.640%
35) Dibromofluoromethane	8.165	113	98459	49.868	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.740%
50) Toluene-d8	10.565	98	352871	48.578	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.160%
62) 4-Bromofluorobenzene	12.847	95	118686	44.849	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.700%
Target Compounds						
16) Acetone	4.448	43	4008	3.679	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084435.D
Acq On : 21 Oct 2024 14:53
Operator : JC\MD
Sample : P4438-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VPB190-HYD-20241016

Quant Time: Oct 22 01:34:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

