

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102219\
 Data File : VN058898.D
 Acq On : 22 Oct 2019 10:18
 Operator : JC/SP
 Sample : VN1022WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1022WBL01

Quant Time: Oct 23 04:53:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 04:07:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	1843285	5.00	ug/l	-0.02
4) 1,4-Difluorobenzene	8.56	114	2513081	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	2256478	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	698708	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.00	65	1161960	5.28	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	105.60%
5) Dibromofluoromethane	7.57	113	891897	5.63	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	112.60%
6) Toluene-d8	10.08	98	1947731	5.44	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	108.80%
7) 4-Bromofluorobenzene	12.40	95	776467	4.43	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	88.60%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102219\
Data File : VN058898.D
Acq On : 22 Oct 2019 10:18
Operator : JC/SP
Sample : VN1022WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1022WBL01

Quant Time: Oct 23 04:53:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 04:07:23 2019
Response via : Initial Calibration

