

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102219\
 Data File : VN058901.D
 Acq On : 22 Oct 2019 11:38
 Operator : JC/SP
 Sample : K5452-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FA19-MMW0421B

Quant Time: Oct 23 04:53:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 04:07:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	1723796	5.00	ug/l	-0.02
4) 1,4-Difluorobenzene	8.56	114	2407947	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	2139528	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	639700	5.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,2-Dichloroethane-d4		8.00	65	1059722	5.15	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	103.00%	
5) Dibromofluoromethane		7.57	113	824053	5.43	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	108.60%	
6) Toluene-d8		10.08	98	1704447	4.97	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	99.40%	
7) 4-Bromofluorobenzene		12.40	95	666594	3.97	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	79.40%#	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102219\
Data File : VN058901.D
Acq On : 22 Oct 2019 11:38
Operator : JC/SP
Sample : K5452-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FA19-MMW0421B

Quant Time: Oct 23 04:53:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 04:07:23 2019
Response via : Initial Calibration

