

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN102219\
Data File : VN058904.D
Acq On : 22 Oct 2019 13:06
Operator : JC/SP
Sample : VSSTDCCC005
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSSTDCCC005

Quant Time: Oct 23 04:54:11 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 04:07:23 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	54	-0.02
2 C	Vinyl Chloride	0.508	0.588	-15.7#	60	0.00
3 S	1,2-Dichloroethane-d4	0.597	0.645	-8.0	56	0.00
4 I	1,4-Difluorobenzene	1.000	1.000	0.0	51	0.00
5 S	Dibromofluoromethane	0.315	0.362	-14.9	56	0.00
6 S	Toluene-d8	0.712	0.735	-3.2	51	0.00
7 S	4-Bromofluorobenzene	0.349	0.292	16.3	43#	0.00
8 I	Chlorobenzene-d5	1.000	1.000	0.0	53	0.00
9 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	40#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1