

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102220\
Data File : VN064254.D
Acq On : 22 Oct 2020 16:56
Operator : JC/MD
Sample : VN1022WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1022WBL01

Quant Time: Oct 23 01:21:47 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N102020W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 22 16:07:21 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	39247	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	225030	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	192374	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.99	65	107324	33.40	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	111.33%#
60) 4-Bromofluorobenzene	12.38	95	97181	28.19	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.97%
63) Toluene-d8	10.06	98	262158	30.10	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.33%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102220\
Data File : VN064254.D
Acq On : 22 Oct 2020 16:56
Operator : JC/MD
Sample : VN1022WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1022WBL01

Quant Time: Oct 23 01:21:47 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N102020W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 22 16:07:21 2020
Response via : Initial Calibration

