

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102220\
 Data File : VN064256.D
 Acq On : 22 Oct 2020 17:43
 Operator : JC/MD
 Sample : VN1022WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1022WBL02

Quant Time: Oct 23 01:26:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N102020W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Oct 22 16:07:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	46271	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	260582	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	226901	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.99	65	127851	33.75	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	112.50%#
60) 4-Bromofluorobenzene	12.38	95	119826	29.47	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	98.23%
63) Toluene-d8	10.06	98	306782	29.86	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.53%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102220\
Data File : VN064256.D
Acq On : 22 Oct 2020 17:43
Operator : JC/MD
Sample : VN1022WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1022WBL02

Quant Time: Oct 23 01:26:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N102020W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 22 16:07:21 2020
Response via : Initial Calibration

