

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102221\
 Data File : VN069134.D
 Acq On : 22 Oct 2021 14:33
 Operator : JC/MD
 Sample : VN1022WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1022WBL01

Quant Time: Oct 23 07:54:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N102221W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 22 13:15:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.662	128	52779	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.968	114	313270	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.747	117	269451	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	127222	30.947	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.167%
60) 4-Bromofluorobenzene	12.734	95	118687	25.689	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	85.633%
63) Toluene-d8	10.443	98	375336	30.831	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.767%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102221\
Data File : VN069134.D
Acq On : 22 Oct 2021 14:33
Operator : JC/MD
Sample : VN1022WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1022WBL01

Quant Time: Oct 23 07:54:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N102221W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 22 13:15:19 2021
Response via : Initial Calibration

