

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102221\
 Data File : VN069144.D
 Acq On : 22 Oct 2021 19:49
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Oct 23 07:56:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N102221W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 22 13:15:19 2021
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	80	0.00
2 M	Dichlorodifluoromethane	1.984	2.119	-6.8	81	0.00
3 M	Chloromethane	2.062	2.206	-7.0	82	0.00
4 M	Vinyl Chloride	2.458	2.554	-3.9	79	0.00
5 M	Bromomethane	1.612	1.679	-4.2	72	0.00
6 M	Chloroethane	1.593	1.634	-2.6	70	0.00
7 M	Trichlorofluoromethane	3.183	3.351	-5.3	79	0.00
8 T	Diethyl Ether	1.127	1.210	-7.4	79	0.00
9	1,1,2-Trichlorotrifluoroeth	1.729	1.791	-3.6	78	0.00
10 M	1,1-Dichloroethene	1.687	1.763	-4.5	80	0.00
11	Methyl Iodide	2.372	2.203	7.1	75	0.00
12	Methyl Acetate	3.213	3.386	-5.4	81	0.00
13 M	Acrolein	0.194	0.175	9.8	72	0.00
14 M	Acrylonitrile	0.912	0.962	-5.5	81	0.00
15 M	Acetone	0.285	0.251	11.9	58	0.00
16 M	Carbon Disulfide	4.534	4.471	1.4	78	0.00
17	Allyl chloride	2.906	2.931	-0.9	78	0.00
18 M	Methylene Chloride	2.005	2.125	-6.0	81	0.00
19 M	trans-1,2-Dichloroethene	1.832	1.874	-2.3	79	0.00
20 T	Diisopropyl ether	6.207	6.311	-1.7	79	0.00
21 M	1,1-Dichloroethane	3.464	3.580	-3.3	79	0.00
22 M	cis-1,2-Dichloroethene	2.196	2.243	-2.1	79	0.00
23 M	tert-Butyl Alcohol	0.374	0.386	-3.2	80	0.00
24 M	Methyl tert-Butyl Ether	6.400	6.638	-3.7	80	0.00
25 M	Chloroform	3.702	3.815	-3.1	79	0.00
26	Cyclohexane	3.191	3.248	-1.8	78	0.00
27 s	1,2-Dichloroethane-d4	2.337	2.374	-1.6	81	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	79	0.00
29	1,1-Dichloropropene	0.472	0.472	0.0	78	0.00
30 M	2-Butanone	0.227	0.225	0.9	73	0.00
31	2,2-Dichloropropane	0.566	0.514	9.2	70	0.00
32 M	1,1,1-Trichloroethane	0.587	0.593	-1.0	78	0.00
33 M	Carbon Tetrachloride	0.499	0.499	0.0	78	0.00
34 M	Benzene	1.410	1.449	-2.8	79	0.00
35	Methacrylonitrile	0.226	0.235	-4.0	84	0.00
36 M	1,2-Dichloroethane	0.509	0.517	-1.6	80	0.00
37 M	Trichloroethene	0.358	0.363	-1.4	79	0.00
38	Methylcyclohexane	0.601	0.602	-0.2	78	0.00
39 M	1,2-Dichloropropane	0.359	0.368	-2.5	78	0.00
40	Dibromomethane	0.251	0.258	-2.8	80	0.00
41 M	Bromodichloromethane	0.502	0.501	0.2	78	0.00
42 M	Vinyl Acetate	0.809	0.813	-0.5	78	0.00
43	Ethyl Acetate	0.442	0.468	-5.9	80	0.00
44	Isopropyl Acetate	0.819	0.814	0.6	76	0.00
45 T	1,4-Dioxane	0.007	0.007#	0.0	81	0.00
46	Methyl methacrylate	0.370	0.351	5.1	79	0.00
47	n-amyl Acetate	0.695	0.666	4.2	78	0.00
48 M	t-1,3-Dichloropropene	0.581	0.548	5.7	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102221\
 Data File : VN069144.D
 Acq On : 22 Oct 2021 19:49
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Oct 23 07:56:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N102221W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 22 13:15:19 2021
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.605	0.582	3.8	76	0.00
50 M	1,1,2-Trichloroethane	0.361	0.370	-2.5	80	0.00
51	Ethyl methacrylate	0.595	0.598	-0.5	79	0.00
52	1,3-Dichloropropane	0.607	0.620	-2.1	79	0.00
53 M	Dibromochloromethane	0.385	0.377	2.1	79	0.00
54 M	1,2-Dibromoethane	0.377	0.382	-1.3	79	0.00
55 M	2-Chloroethyl vinyl ether	0.189	0.185	2.1	82	0.00
56 M	Bromoform	0.278	0.263	5.4	79	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
58 M	4-Methyl-2-Pentanone	0.506	0.543	-7.3	80	0.00
59 M	2-Hexanone	0.382	0.393	-2.9	75	0.00
60 S	4-Bromofluorobenzene	0.514	0.501	2.5	78	0.00
61 M	Tetrachloroethene	0.326	0.361	-10.7	84	0.00
62 M	Toluene	1.747	1.821	-4.2	78	0.00
63 S	Toluene-d8	1.355	1.387	-2.4	79	0.00
64 M	Chlorobenzene	1.060	1.093	-3.1	79	0.00
65	1,1,1,2-Tetrachloroethane	0.394	0.409	-3.8	79	0.00
66 M	Ethyl Benzene	1.965	2.030	-3.3	78	0.00
67 M	m/p-Xylenes	0.739	0.762	-3.1	78	0.00
68 M	o-Xylene	0.740	0.763	-3.1	79	0.00
69 M	Styrene	1.226	1.234	-0.7	79	0.00
70	Isopropylbenzene	1.939	1.987	-2.5	79	0.00
71 M	1,1,2,2-Tetrachloroethane	0.568	0.592	-4.2	77	0.00
72	1,2,3-Trichloropropane	0.488	0.504	-3.3	85	0.00
73	Bromobenzene	0.440	0.449	-2.0	79	0.00
74	n-propylbenzene	2.227	2.264	-1.7	79	0.00
75	2-Chlorotoluene	1.364	1.394	-2.2	78	0.00
76	1,3,5-Trimethylbenzene	1.616	1.673	-3.5	79	0.00
77	t-1,4-Dichloro-2-butene	0.191	0.170	11.0	72	0.00
78	4-Chlorotoluene	1.337	1.326	0.8	78	0.00
79	tert-butylbenzene	1.419	1.462	-3.0	78	0.00
80	1,2,4-Trimethylbenzene	1.587	1.620	-2.1	78	0.00
81	sec-Butylbenzene	1.987	2.020	-1.7	79	0.00
82	p-Isopropyltoluene	1.621	1.604	1.0	78	0.00
83 M	1,3-Dichlorobenzene	0.774	0.773	0.1	79	0.00
84 M	1,4-Dichlorobenzene	0.754	0.753	0.1	80	0.00
85	n-Butylbenzene	1.355	1.279	5.6	77	0.00
86 T	Hexachloroethane	0.303	0.294	3.0	77	0.00
87 M	1,2-Dichlorobenzene	0.739	0.748	-1.2	79	0.00
88	1,2-Dibromo-3-Chloropropane	0.106	0.107	-0.9	79	0.00
89	1,2,4-Trichlorobenzene	0.373	0.361	3.2	81	0.00
90	Hexachlorobutadiene	0.195	0.197	-1.0	79	0.00
91 M	Naphthalene	1.078	1.088	-0.9	87	0.00
92	1,2,3-Trichlorobenzene	0.356	0.358	-0.6	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0