

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102224\
 Data File : VN084452.D
 Acq On : 22 Oct 2024 13:38
 Operator : JC\MD
 Sample : VN1022WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1022WBL01

Quant Time: Oct 23 01:26:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	195991	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	325943	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	276681	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	106919	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	132772	45.690	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.380%
35) Dibromofluoromethane	8.165	113	107332	49.949	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.900%
50) Toluene-d8	10.565	98	389764	49.301	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.600%
62) 4-Bromofluorobenzene	12.847	95	129594	44.996	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102224\
Data File : VN084452.D
Acq On : 22 Oct 2024 13:38
Operator : JC\MD
Sample : VN1022WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1022WBL01

Quant Time: Oct 23 01:26:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

