

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102224\
 Data File : VN084459.D
 Acq On : 22 Oct 2024 17:09
 Operator : JC\MD
 Sample : P4475-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-190-GW-278-280

Quant Time: Oct 23 01:31:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	157977	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	281977	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245983	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	100831	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	117218	50.044	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.080%	
35) Dibromofluoromethane	8.165	113	95280	51.254	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.500%	
50) Toluene-d8	10.565	98	341267	49.898	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.800%	
62) 4-Bromofluorobenzene	12.847	95	115495	46.353	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 92.700%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102224\
Data File : VN084459.D
Acq On : 22 Oct 2024 17:09
Operator : JC\MD
Sample : P4475-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-190-GW-278-280

Quant Time: Oct 23 01:31:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

