

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102224\
 Data File : VN084462.D
 Acq On : 22 Oct 2024 18:21
 Operator : JC\MD
 Sample : P4397-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WB-301-BOT

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/23/2024
 Supervised By :Mahesh Dadoda 10/23/2024

Quant Time: Oct 23 01:32:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168250	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	297819	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	273036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103880	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	123682	49.580	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.160%		
35) Dibromofluoromethane	8.165	113	98197	50.014	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.020%		
50) Toluene-d8	10.565	98	376295	52.092	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	104.180%		
62) 4-Bromofluorobenzene	12.847	95	127222	48.344	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.680%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	26114	24.368	ug/l	95
20) Methylene Chloride	5.265	84	3526m	1.619	ug/l	
43) Isopropyl Acetate	8.688	43	141080	24.618	ug/l #	88
95) Naphthalene	15.647	128	10247	1.784	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102224\
Data File : VN084462.D
Acq On : 22 Oct 2024 18:21
Operator : JC\MD
Sample : P4397-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WB-301-BOT

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/23/2024
Supervised By :Mahesh Dadoda 10/23/2024

Quant Time: Oct 23 01:32:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

