

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN102224\  
 Data File : VN084463.D  
 Acq On : 22 Oct 2024 18:46  
 Operator : JC\MD  
 Sample : P4460-04  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 WB-303-BOT

Quant Time: Oct 23 01:32:43 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N093024W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Oct 01 07:11:01 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 8.224  | 168   | 163501   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.100  | 114   | 298176   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.865 | 117   | 267301   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.794 | 152   | 97626    | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.577  | 65    | 120367   | 49.652   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =      | 99.300%  |
| 35) Dibromofluoromethane    | 8.171  | 113   | 97567    | 49.633   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 99.260%  |
| 50) Toluene-d8              | 10.565 | 98    | 365187   | 50.494   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =      | 100.980% |
| 62) 4-Bromofluorobenzene    | 12.847 | 95    | 119941   | 45.522   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | =      | 91.040%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 4.442  | 43    | 27887    | 26.779   | ug/l   | 94       |
| 20) Methylene Chloride      | 5.283  | 84    | 4490     | 2.121    | ug/l # | 66       |
| 43) Isopropyl Acetate       | 8.688  | 43    | 154605   | 26.945   | ug/l # | 83       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN102224\  
Data File : VN084463.D  
Acq On : 22 Oct 2024 18:46  
Operator : JC\MD  
Sample : P4460-04  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
WB-303-BOT

Quant Time: Oct 23 01:32:43 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N093024W.M  
Quant Title : SW846 8260  
QLast Update : Tue Oct 01 07:11:01 2024  
Response via : Initial Calibration

