

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102318\
 Data File : VN051966.D
 Acq On : 23 Oct 2018 16:13
 Operator : MD\SY
 Sample : I.BLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 I.BLK

Quant Time: Oct 24 07:28:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 06:55:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1152825	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1676898	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1520954	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	721746	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	575623	56.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.12%	
35) Dibromofluoromethane	7.59	113	612228	57.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.44%	
50) Toluene-d8	10.09	98	2256610	58.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	116.72%	
62) 4-Bromofluorobenzene	12.40	95	725654	54.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.20%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102318\
Data File : VN051966.D
Acq On : 23 Oct 2018 16:13
Operator : MD\SY
Sample : I.BLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
I.BLK

Quant Time: Oct 24 07:28:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 06:55:59 2018
Response via : Initial Calibration

