

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102318\
 Data File : VN051968.D
 Acq On : 23 Oct 2018 17:46
 Operator : MD\SY
 Sample : VN1023WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1023WBL01

Quant Time: Oct 24 07:37:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N101218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 12 10:23:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	154987	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	890453	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	753517	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	289321	30.68	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.27%
60) 4-Bromofluorobenzene	12.40	95	296545	27.29	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	90.97%
63) Toluene-d8	10.09	98	984511	30.09	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.30%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102318\
Data File : VN051968.D
Acq On : 23 Oct 2018 17:46
Operator : MD\SY
Sample : VN1023WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN1023WBL01

Quant Time: Oct 24 07:37:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N101218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 12 10:23:44 2018
Response via : Initial Calibration

