

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102318\
 Data File : VN051970.D
 Acq On : 23 Oct 2018 18:39
 Operator : MD\SY
 Sample : J5643-01 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 001M-WILLETS-PT-BLVD-(SEP)

Quant Time: Oct 24 07:44:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N101218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 12 10:23:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	153758	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	829576	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	664177	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	281319	30.07	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.23%
60) 4-Bromofluorobenzene	12.40	95	236225	24.66	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	82.20%
63) Toluene-d8	10.09	98	906428	31.43	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	104.77%

Target Compounds

					Ovalue
12) Methyl Acetate	4.33	43	13654	3.106 ug/l	92
15) Acetone	3.82	58	17442	35.913 ug/l	96
25) Chloroform	7.37	83	97800	6.303 ug/l	99
41) Bromodichloromethane	9.40	83	7791	0.682 ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102318\
Data File : VN051970.D
Acq On : 23 Oct 2018 18:39
Operator : MD\SY
Sample : J5643-01 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
001M-WILLETS-PT-BLVD-(SEP)

Quant Time: Oct 24 07:44:49 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N101218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 12 10:23:44 2018
Response via : Initial Calibration

