

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102320\
 Data File : VN064301.D
 Acq On : 23 Oct 2020 21:39
 Operator : JC/MD
 Sample : L4486-08 100X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1814

Quant Time: Oct 26 02:51:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 21 12:48:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	239504	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	415883	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	417645	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	192599	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	188794	53.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.16%	
35) Dibromofluoromethane	7.55	113	140060	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
50) Toluene-d8	10.06	98	527821	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
62) 4-Bromofluorobenzene	12.38	95	230957	56.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.60%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.02	105	8727	0.706	ug/l	97
86) p-Isopropyltoluene	13.26	119	2622	0.224	ug/l	83
95) Naphthalene	15.10	128	12912	1.270	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102320\
Data File : VN064301.D
Acq On : 23 Oct 2020 21:39
Operator : JC/MD
Sample : L4486-08 100X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1814

Quant Time: Oct 26 02:51:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102120W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 21 12:48:40 2020
Response via : Initial Calibration

