

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
 Data File : VN079635.D
 Acq On : 23 Oct 2023 09:28
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Oct 23 23:44:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	265098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	449732	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	412239	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	185298	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	211645	49.239	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.480%
35) Dibromofluoromethane	8.171	113	169793	52.536	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.080%
50) Toluene-d8	10.571	98	623869	53.349	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.700%
62) 4-Bromofluorobenzene	12.847	95	204202	54.412	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.820%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.130	85	157110	41.359	ug/l	99
3) Chloromethane	2.365	50	191267	40.353	ug/l	99
4) Vinyl Chloride	2.518	62	210997	42.797	ug/l	96
5) Bromomethane	2.954	94	117931	39.957	ug/l	100
6) Chloroethane	3.124	64	141192	37.986	ug/l	98
7) Trichlorofluoromethane	3.501	101	265889	45.677	ug/l	100
8) Diethyl Ether	3.965	74	94434	45.505	ug/l	63
9) 1,1,2-Trichlorotrifluo...	4.377	101	152704	44.890	ug/l	100
10) Methyl Iodide	4.595	142	175078	59.311	ug/l	96
11) Tert butyl alcohol	5.524	59	111184	171.863	ug/l	98
12) 1,1-Dichloroethene	4.342	96	141690	45.237	ug/l	82
13) Acrolein	4.177	56	103265	160.307	ug/l	98
14) Allyl chloride	5.030	41	224400	44.396	ug/l #	90
15) Acrylonitrile	5.718	53	405993	216.463	ug/l	99
16) Acetone	4.424	43	484309	284.073	ug/l	92
17) Carbon Disulfide	4.718	76	398783	40.817	ug/l	99
18) Methyl Acetate	5.024	43	327909	39.297	ug/l #	86
19) Methyl tert-butyl Ether	5.800	73	485203	47.677	ug/l	94
20) Methylene Chloride	5.283	84	176386	47.212	ug/l #	83
21) trans-1,2-Dichloroethene	5.789	96	162830	44.510	ug/l #	81
22) Diisopropyl ether	6.677	45	554213	48.898	ug/l #	91
23) Vinyl Acetate	6.606	43	1472460	230.959	ug/l #	87
24) 1,1-Dichloroethane	6.571	63	325888	46.981	ug/l	97
25) 2-Butanone	7.489	43	591223	233.648	ug/l #	82
26) 2,2-Dichloropropane	7.494	77	268482	50.646	ug/l	98
27) cis-1,2-Dichloroethene	7.489	96	190718	47.989	ug/l	86
28) Bromochloromethane	7.818	49	164819	49.939	ug/l #	73
29) Tetrahydrofuran	7.841	42	348493	212.947	ug/l #	75
30) Chloroform	7.965	83	331579	47.310	ug/l	93
31) Cyclohexane	8.259	56	259186	42.510	ug/l	88
32) 1,1,1-Trichloroethane	8.171	97	279477	47.384	ug/l	94
36) 1,1-Dichloropropene	8.371	75	244817	50.471	ug/l	95
37) Ethyl Acetate	7.565	43	219513	45.480	ug/l #	90
38) Carbon Tetrachloride	8.365	117	237857	48.982	ug/l	98
39) Methylcyclohexane	9.606	83	229124	49.165	ug/l #	85
40) Benzene	8.606	78	728217	50.551	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
 Data File : VN079635.D
 Acq On : 23 Oct 2023 09:28
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Oct 23 23:44:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	115395	45.476	ug/l #	85
42) 1,2-Dichloroethane	8.671	62	257871	49.264	ug/l	99
43) Isopropyl Acetate	8.688	43	359428	42.506	ug/l #	91
44) Trichloroethene	9.353	130	169376	49.954	ug/l	98
45) 1,2-Dichloropropane	9.624	63	194794	51.321	ug/l	97
46) Dibromomethane	9.712	93	127923	52.038	ug/l	97
47) Bromodichloromethane	9.888	83	264926	52.391	ug/l	98
48) Methyl methacrylate	9.683	41	170726	47.116	ug/l #	84
49) 1,4-Dioxane	9.694	88	55269	828.070	ug/l #	82
51) 4-Methyl-2-Pentanone	10.447	43	1124107	235.493	ug/l	89
52) Toluene	10.635	92	447989	52.831	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	269969	52.978	ug/l	98
54) cis-1,3-Dichloropropene	10.318	75	297781	53.381	ug/l #	89
55) 1,1,2-Trichloroethane	11.018	97	175518	51.557	ug/l	95
56) Ethyl methacrylate	10.877	69	250810	53.615	ug/l #	83
57) 1,3-Dichloropropane	11.165	76	311438	52.980	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.165	63	483395	269.826	ug/l	89
59) 2-Hexanone	11.200	43	845798	245.548	ug/l	88
60) Dibromochloromethane	11.359	129	184719	53.989	ug/l	99
61) 1,2-Dibromoethane	11.471	107	168785	49.959	ug/l	98
64) Tetrachloroethene	11.106	164	137729	48.567	ug/l	97
65) Chlorobenzene	11.894	112	453487	49.396	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.965	131	175898	52.328	ug/l	99
67) Ethyl Benzene	11.965	91	844632	52.932	ug/l	96
68) m/p-Xylenes	12.071	106	640459	110.464	ug/l	95
69) o-Xylene	12.400	106	300430	52.689	ug/l	90
70) Styrene	12.412	104	514032	57.074	ug/l	98
71) Bromoform	12.582	173	123349	50.138	ug/l #	99
73) Isopropylbenzene	12.700	105	780571	45.280	ug/l	99
74) N-amyl acetate	12.494	43	293859	41.581	ug/l #	89
75) 1,1,2,2-Tetrachloroethane	12.941	83	249847	44.776	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	304252	56.948	ug/l	72
77) Bromobenzene	12.982	156	185539	43.895	ug/l	98
78) n-propylbenzene	13.035	91	935613	47.970	ug/l	98
79) 2-Chlorotoluene	13.129	91	560766	44.931	ug/l	99
80) 1,3,5-Trimethylbenzene	13.176	105	651243	47.908	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.741	75	73719	40.036	ug/l #	86
82) 4-Chlorotoluene	13.223	91	549727	46.288	ug/l	100
83) tert-Butylbenzene	13.441	119	522552	46.919	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	649518	50.067	ug/l	99
85) sec-Butylbenzene	13.618	105	734869	47.305	ug/l	100
86) p-Isopropyltoluene	13.729	119	614253	51.542	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	327807	48.896	ug/l	97
88) 1,4-Dichlorobenzene	13.812	146	314696	45.460	ug/l	99
89) n-Butylbenzene	14.059	91	496489	52.170	ug/l	98
90) Hexachloroethane	14.335	117	118915	49.132	ug/l	88
91) 1,2-Dichlorobenzene	14.106	146	310949	47.549	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.723	75	37792	37.913	ug/l	96
93) 1,2,4-Trichlorobenzene	15.394	180	101877	50.300	ug/l	99
94) Hexachlorobutadiene	15.500	225	69737	44.209	ug/l	99
95) Naphthalene	15.641	128	250004	40.422	ug/l	98
96) 1,2,3-Trichlorobenzene	15.847	180	103717	42.912	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
Data File : VN079635.D
Acq On : 23 Oct 2023 09:28
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :


Quant Time: Oct 23 23:44:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 23 23:44:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:168 Resp: 265098
Ion Ratio Lower Upper
168 100
99 56.4 53.0 79.6



#2
Dichlorodifluoromethane
Concen: 41.359 ug/l
RT: 2.130 min Scan# 30
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion: 85 Resp: 157110
Ion Ratio Lower Upper
85 100
87 32.2 15.8 47.4









#3
Chloromethane
Concen: 40.353 ug/l
RT: 2.365 min Scan# 70
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 50 Resp: 191267
Ion Ratio Lower Upper
50 100
52 32.3 26.5 39.7



#4
Vinyl Chloride
Concen: 42.797 ug/l
RT: 2.518 min Scan# 96
Delta R.T. 0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion: 62 Resp: 210997
Ion Ratio Lower Upper
62 100
64 30.5 26.2 39.2









#5
Bromomethane
Concen: 39.957 ug/l
RT: 2.954 min Scan# 11
Delta R.T. 0.018 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 117931
Ion Ratio Lower Upper
94 100
96 91.5 73.1 109.7



#6
Chloroethane
Concen: 37.986 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.012 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion: 64 Resp: 141192
Ion Ratio Lower Upper
64 100
66 30.5 25.4 38.2









#7
 Trichlorofluoromethane
 Concen: 45.677 ug/l
 RT: 3.501 min Scan# 261
 Delta R.T. 0.012 min
 Lab File: VN079635.D
 Acq: 23 Oct 2023 09:28




Instrument :
 MSVOA_N
 ClientSampleId :

Tgt Ion:101 Resp: 265889
 Ion Ratio Lower Upper
 101 100
 103 63.3 50.7 76.1



#8
 Diethyl Ether
 Concen: 45.505 ug/l
 RT: 3.965 min Scan# 342
 Delta R.T. 0.000 min
 Lab File: VN079635.D
 Acq: 23 Oct 2023 09:28

Tgt Ion: 74 Resp: 94434
 Ion Ratio Lower Upper
 74 100
 45 98.8 34.4 103.3








#9
1,1,2-Trichlorotrifluoroethane
Concen: 44.890 ug/l
RT: 4.377 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion:101 Resp: 152704
Ion Ratio Lower Upper
101 100
85 46.0 36.6 55.0
151 71.7 57.6 86.4



#10
Methyl Iodide
Concen: 59.311 ug/l
RT: 4.595 min Scan# 449
Delta R.T. 0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion:142 Resp: 175078
Ion Ratio Lower Upper
142 100
127 45.1 37.6 56.4
141 16.8 11.4 17.0









#11
Tert butyl alcohol
Concen: 171.863 ug/l
RT: 5.524 min Scan# 608
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 59 Resp: 111184
Ion Ratio Lower Upper
59 100
57 11.1 8.2 12.4



#12
1,1-Dichloroethene
Concen: 45.237 ug/l
RT: 4.342 min Scan# 406
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28


Tgt Ion: 96 Resp: 141690
Ion Ratio Lower Upper
96 100
61 168.8 112.7 169.1
98 57.3 51.2 76.8








#13
Acrolein
Concen: 160.307 ug/l
RT: 4.177 min Scan# 31
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 56 Resp: 103265
Ion Ratio Lower Upper
56 100
55 69.8 54.5 81.7



#14
Allyl chloride
Concen: 44.396 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28


Tgt Ion: 41 Resp: 224400
Ion Ratio Lower Upper
41 100
39 78.7 58.8 88.2
76 38.2 39.1 58.7#








#15
Acrylonitrile
Concen: 216.463 ug/l
RT: 5.718 min Scan# 641
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 53 Resp: 405993
Ion Ratio Lower Upper
53 100
52 82.0 65.3 97.9
51 36.4 28.9 43.3



#16
Acetone
Concen: 284.073 ug/l
RT: 4.424 min Scan# 420
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion: 43 Resp: 484309
Ion Ratio Lower Upper
43 100
58 30.3 28.0 42.0









#17
Carbon Disulfide
Concen: 40.817 ug/l
RT: 4.718 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 398783
Ion Ratio Lower Upper
76 100
78 9.7 7.4 11.0



#18
Methyl Acetate
Concen: 39.297 ug/l
RT: 5.024 min Scan# 522
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion: 43 Resp: 327909
Ion Ratio Lower Upper
43 100
74 24.9 26.2 39.4#









#19
Methyl tert-butyl Ether
Concen: 47.677 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 73 Resp: 485203
Ion Ratio Lower Upper
73 100
57 22.5 15.6 23.4



#20
Methylene Chloride
Concen: 47.212 ug/l
RT: 5.283 min Scan# 566
Delta R.T. -0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28


Tgt Ion: 84 Resp: 176386
Ion Ratio Lower Upper
84 100
49 130.7 79.7 119.5#
51 35.9 25.2 37.8
86 65.9 53.3 79.9








#21
trans-1,2-Dichloroethene
Concen: 44.510 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion: 96 Resp: 162830
Ion Ratio Lower Upper
96 100
61 141.3 90.1 135.1#
98 61.8 46.4 69.6



#22
Diisopropyl ether
Concen: 48.898 ug/l
RT: 6.677 min Scan# 803
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28



Tgt Ion: 45 Resp: 554213
Ion Ratio Lower Upper
45 100
43 54.8 45.5 68.3
87 28.0 31.8 47.8#
59 13.5 11.6 17.4









#23
 Vinyl Acetate
 Concen: 230.959 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.000 min
 Lab File: VN079635.D
 Acq: 23 Oct 2023 09:28




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 43 Resp: 1472460
 Ion Ratio Lower Upper
 43 100
 86 10.8 13.0 19.6#



#24
 1,1-Dichloroethane
 Concen: 46.981 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. -0.006 min
 Lab File: VN079635.D
 Acq: 23 Oct 2023 09:28

Tgt Ion: 63 Resp: 325888
 Ion Ratio Lower Upper
 63 100
 98 6.7 3.9 11.7
 100 4.1 2.5 7.4









#25
2-Butanone
Concen: 233.648 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.001 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 43 Resp: 591223
Ion Ratio Lower Upper
43 100
72 26.9 30.2 45.2#



#26
2,2-Dichloropropane
Concen: 50.646 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28


Tgt Ion: 77 Resp: 268482
Ion Ratio Lower Upper
77 100
97 21.6 11.2 33.6








#27
 cis-1,2-Dichloroethene
 Concen: 47.989 ug/l
 RT: 7.489 min Scan# 941
 Delta R.T. -0.006 min
 Lab File: VN079635.D
 Acq: 23 Oct 2023 09:28

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion: 96 Resp: 190718
 Ion Ratio Lower Upper
 96 100
 61 152.2 0.0 255.2
 98 64.5 0.0 128.8



#28
 Bromochloromethane
 Concen: 49.939 ug/l
 RT: 7.818 min Scan# 997
 Delta R.T. 0.000 min
 Lab File: VN079635.D
 Acq: 23 Oct 2023 09:28

Tgt Ion: 49 Resp: 164819
 Ion Ratio Lower Upper
 49 100
 129 2.1 0.0 4.8
 130 67.6 75.3 112.9#





#29

Tetrahydrofuran

Concen: 212.947 ug/l

RT: 7.841 min Scan# 1001

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 42 Resp: 348493

Ion Ratio Lower Upper

42 100

72 43.4 49.8 74.6#

71 40.0 47.4 71.2#



#30

Chloroform

Concen: 47.310 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.006 min

Lab File: VN079635.D




Acq: 23 Oct 2023 09:28


Tgt Ion: 83 Resp: 331579

Ion Ratio Lower Upper

83 100

85 63.3 55.4 83.0





#31

Cyclohexane

Concen: 42.510 ug/l

RT: 8.259 min Scan# 1072

Delta R.T. -0.006 min


Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 56 Resp: 259186

Ion Ratio Lower Upper

56 100

69 29.4 26.8 40.2

84 88.2 81.8 122.6



#32

1,1,1-Trichloroethane

Concen: 47.384 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28




Tgt Ion: 97 Resp: 279477


Ion Ratio Lower Upper

97 100

99 64.4 50.3 75.5

61 51.4 34.7 52.1





#33

1,2-Dichloroethane-d4

Concen: 49.239 ug/l

RT: 8.583 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 65 Resp: 211645

Ion Ratio Lower Upper

65 100

67 52.2 0.0 106.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28




Tgt Ion: 114 Resp: 449732

Ion Ratio Lower Upper

114 100

63 20.4 0.0 36.4


88 15.8 0.0 30.6








#35
Dibromofluoromethane
Concen: 52.536 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:113 Resp: 169793
Ion Ratio Lower Upper
113 100
111 101.6 82.6 123.8
192 18.1 12.7 19.1



#36
1,1-Dichloropropene
Concen: 50.471 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion: 75 Resp: 244817
Ion Ratio Lower Upper
75 100
110 33.2 19.1 57.3
77 30.3 25.0 37.4









#37
Ethyl Acetate
Concen: 45.480 ug/l
RT: 7.565 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 219513
Ion Ratio Lower Upper
43 100
61 12.7 13.2 19.8#
70 9.5 11.3 16.9#



#38
Carbon Tetrachloride
Concen: 48.982 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion: 117 Resp: 237857
Ion Ratio Lower Upper
117 100
119 94.8 77.3 115.9
121 30.8 23.3 34.9









#39
Methylcyclohexane
Concen: 49.165 ug/l
RT: 9.606 min Scan# 1131
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 83 Resp: 229124
Ion Ratio Lower Upper
83 100
55 78.4 49.1 73.7#
98 46.7 40.6 61.0



#40
Benzene
Concen: 50.551 ug/l
RT: 8.606 min Scan# 1131
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion: 78 Resp: 728217
Ion Ratio Lower Upper
78 100
77 22.9 19.0 28.6








#41
Methacrylonitrile
Concen: 45.476 ug/l
RT: 7.783 min Scan# 991
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	41	Resp:	115395
Ion	Ratio	Lower	Upper
41	100		
39	57.6	42.6	64.0
67	74.1	77.5	116.3
52	33.9	28.9	43.3



#42
1,2-Dichloroethane
Concen: 49.264 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion:	62	Resp:	257871
Ion	Ratio	Lower	Upper
62	100		
98	9.5	0.0	19.8









#43
Isopropyl Acetate
Concen: 42.506 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 359428
Ion Ratio Lower Upper
43 100
61 26.4 24.7 37.1
87 13.1 14.6 21.8#



#44
Trichloroethene
Concen: 49.954 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion:130 Resp: 169376
Ion Ratio Lower Upper
130 100
95 103.8 0.0 204.6





#45

1,2-Dichloropropane

Concen: 51.321 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 63 Resp: 194794

Ion Ratio Lower Upper

63 100

65 31.5 23.9 35.9



#46

Dibromomethane

Concen: 52.038 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28




Tgt Ion: 93 Resp: 127923


Ion Ratio Lower Upper

93 100

95 80.9 63.8 95.8

174 86.0 65.6 98.4





#47

Bromodichloromethane

Concen: 52.391 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.006 min


Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 83 Resp: 264926

Ion Ratio Lower Upper

83 100

85 63.6 49.6 74.4

127 8.4 7.0 10.6



#48

Methyl methacrylate

Concen: 47.116 ug/l

RT: 9.683 min Scan# 1314

Delta R.T. -0.006 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28




Tgt Ion: 41 Resp: 170726


Ion Ratio Lower Upper

41 100

69 88.5 90.4 135.6#

39 63.6 54.6 81.8









#49
1,4-Dioxane
Concen: 828.070 ug/l
RT: 9.694 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 88 Resp: 55269
Ion Ratio Lower Upper
88 100
43 29.6 17.0 25.6#
58 71.9 47.0 70.4#



#50
Toluene-d8
Concen: 53.349 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion: 98 Resp: 623869
Ion Ratio Lower Upper
98 100
100 64.6 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 235.493 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 1124107
Ion Ratio Lower Upper
43 100
58 38.3 36.6 54.8



#52
Toluene
Concen: 52.831 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion: 92 Resp: 447989
Ion Ratio Lower Upper
92 100
91 173.7 137.4 206.0









#53
 t-1,3-Dichloropropene
 Concen: 52.978 ug/l
 RT: 10.835 min Scan# 1511
 Delta R.T. -0.006 min
 Lab File: VN079635.D
 Acq: 23 Oct 2023 09:28




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 75 Resp: 269969
 Ion Ratio Lower Upper
 75 100
 77 30.7 25.4 38.0



#54
 cis-1,3-Dichloropropene
 Concen: 53.381 ug/l
 RT: 10.318 min Scan# 1422
 Delta R.T. 0.000 min
 Lab File: VN079635.D
 Acq: 23 Oct 2023 09:28

Tgt Ion: 75 Resp: 297781
 Ion Ratio Lower Upper
 75 100
 77 29.7 26.1 39.1
 39 48.8 31.6 47.4#





#55

1,1,2-Trichloroethane

Concen: 51.557 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. -0.006 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion: 97 Resp: 175518





Ion Ratio Lower Upper

97 100

83 89.2 65.1 97.7

85 55.7 43.4 65.2

99 59.1 49.2 73.8



#56

Ethyl methacrylate

Concen: 53.615 ug/l

RT: 10.877 min Scan# 1517

Delta R.T. -0.005 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28




Tgt Ion: 69 Resp: 250810


Ion Ratio Lower Upper

69 100

41 66.4 39.6 59.4#

39 39.7 28.6 43.0





#57

1,3-Dichloropropane

Concen: 52.980 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. -0.006 min


Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 76 Resp: 311438

Ion Ratio Lower Upper

76 100

78 32.0 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 269.826 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. -0.000 min

Lab File: VN079635.D




Acq: 23 Oct 2023 09:28


Tgt Ion: 63 Resp: 483395

Ion Ratio Lower Upper

63 100

106 26.5 26.2 39.4









#59
2-Hexanone
Concen: 245.548 ug/l
RT: 11.200 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 845798
Ion Ratio Lower Upper
43 100
58 54.0 31.7 95.1



#60
Dibromochloromethane
Concen: 53.989 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion:129 Resp: 184719
Ion Ratio Lower Upper
129 100
127 78.4 38.6 116.0





#61

1,2-Dibromoethane

Concen: 49.959 ug/l

RT: 11.471 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:107 Resp: 168785

Ion Ratio Lower Upper

107 100

109 93.6 76.6 114.8



#62

4-Bromofluorobenzene

Concen: 54.412 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. -0.006 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28




Tgt Ion: 95 Resp: 204202


Ion Ratio Lower Upper

95 100

174 77.1 0.0 143.8

176 73.5 0.0 135.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:117 Resp: 412239
Ion Ratio Lower Upper
117 100
82 56.5 44.0 66.0
119 32.0 25.6 38.4



#64
Tetrachloroethene
Concen: 48.567 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion:164 Resp: 137729
Ion Ratio Lower Upper
164 100
166 129.6 101.4 152.0
129 98.7 82.4 123.6
131 94.2 78.2 117.4





#65

Chlorobenzene

Concen: 49.396 ug/l

RT: 11.894 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:112 Resp: 453487

Ion Ratio Lower Upper

112 100

114 31.8 25.0 37.4



#66

1,1,1,2-Tetrachloroethane

Concen: 52.328 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28




Tgt Ion:131 Resp: 175898


Ion Ratio Lower Upper

131 100

133 92.7 46.4 139.1

119 65.4 34.1 102.2









#67
Ethyl Benzene
Concen: 52.932 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 844632
Ion Ratio Lower Upper
91 100
106 29.7 25.6 38.4



#68
m/p-Xylenes
Concen: 110.464 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. -0.005 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion:106 Resp: 640459
Ion Ratio Lower Upper
106 100
91 207.4 159.6 239.4









#69
o-Xylene
Concen: 52.689 ug/l
RT: 12.400 min Scan# 1777
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 300430
Ion Ratio Lower Upper
106 100
91 225.6 105.2 315.5



#70
Styrene
Concen: 57.074 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion:104 Resp: 514032
Ion Ratio Lower Upper
104 100
78 50.7 39.5 59.3
103 55.5 43.2 64.8





#71

Bromoform

Concen: 50.138 ug/l

RT: 12.582 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:173 Resp: 123349

Ion Ratio Lower Upper

173 100

175 49.3 24.3 73.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28




Tgt Ion:152 Resp: 185298


Ion Ratio Lower Upper

152 100

115 62.7 32.3 96.8

150 176.1 0.0 355.8









#73
Isopropylbenzene
Concen: 45.280 ug/l
RT: 12.700 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 780571
Ion Ratio Lower Upper
105 100
120 25.4 13.1 39.1



#74
N-ethyl acetate
Concen: 41.581 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion: 43 Resp: 293859
Ion Ratio Lower Upper
43 100
70 42.5 43.4 65.2#
55 26.4 21.4 32.0
61 23.1 22.8 34.2





#75

1,1,2,2-Tetrachloroethane

Concen: 44.776 ug/l

RT: 12.941 min Scan# 1868

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 83 Resp: 249847

Ion Ratio Lower Upper

83 100

131 9.9 5.0 15.0

85 64.5 31.9 95.7



#76

1,2,3-Trichloropropane

Concen: 56.948 ug/l

RT: 12.994 min Scan# 1877

Delta R.T. -0.006 min

Lab File: VN079635.D




Acq: 23 Oct 2023 09:28


Tgt Ion: 75 Resp: 304252

Ion Ratio Lower Upper

75 100

77 130.8 84.5 253.5





#77

Bromobenzene

Concen: 43.895 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:156 Resp: 185539

Ion Ratio Lower Upper

156 100

77 214.4 107.7 323.3

158 93.2 48.4 145.3



#78

n-propylbenzene

Concen: 47.970 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.006 min

Lab File: VN079635.D




Acq: 23 Oct 2023 09:28


Tgt Ion: 91 Resp: 935613

Ion Ratio Lower Upper

91 100

120 21.8 11.5 34.4









#79
2-Chlorotoluene
Concen: 44.931 ug/l
RT: 13.129 min Scan# 1900
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 560766
Ion Ratio Lower Upper
91 100
126 32.2 16.3 48.8



#80
1,3,5-Trimethylbenzene
Concen: 47.908 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion: 105 Resp: 651243
Ion Ratio Lower Upper
105 100
120 47.7 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 40.036 ug/l
RT: 12.741 min Scan# 1834
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 73719
Ion Ratio Lower Upper
75 100
53 102.1 67.4 101.2#
89 46.5 38.8 58.2



#82
4-Chlorotoluene
Concen: 46.288 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28

Tgt Ion: 91 Resp: 549727
Ion Ratio Lower Upper
91 100
126 33.1 16.7 50.0





#83

tert-Butylbenzene

Concen: 46.919 ug/l

RT: 13.441 min Scan# 1953

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:119 Resp: 522552

Ion Ratio Lower Upper

119 100

91 68.6 33.9 101.7

134 27.5 13.2 39.6



#84

1,2,4-Trimethylbenzene

Concen: 50.067 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.006 min

Lab File: VN079635.D




Acq: 23 Oct 2023 09:28


Tgt Ion:105 Resp: 649518

Ion Ratio Lower Upper

105 100

120 44.6 22.0 66.0









#85
 sec-Butylbenzene
 Concen: 47.305 ug/l
 RT: 13.618 min Scan# 1984
 Delta R.T. -0.006 min
 Lab File: VN079635.D
 Acq: 23 Oct 2023 09:28




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:105 Resp: 734869
 Ion Ratio Lower Upper
 105 100
 134 19.7 9.8 29.3



#86
 p-Isopropyltoluene
 Concen: 51.542 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.006 min
 Lab File: VN079635.D
 Acq: 23 Oct 2023 09:28


Tgt Ion:119 Resp: 614253
 Ion Ratio Lower Upper
 119 100
 134 24.8 12.6 37.8
 91 23.8 12.4 37.2








#87
1,3-Dichlorobenzene
Concen: 48.896 ug/l
RT: 13.735 min Scan# 2003
Delta R.T. 0.000 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:146 Resp: 327807
Ion Ratio Lower Upper
146 100
111 41.1 22.1 66.3
148 62.9 32.0 96.0




#88
1,4-Dichlorobenzene
Concen: 45.460 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. -0.006 min
Lab File: VN079635.D
Acq: 23 Oct 2023 09:28



Tgt Ion:146 Resp: 314696
Ion Ratio Lower Upper
146 100
111 41.1 20.6 61.8
148 63.8 31.6 94.8





#89

n-Butylbenzene

Concen: 52.170 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. 0.000 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 91 Resp: 496489

Ion Ratio Lower Upper

91 100

92 51.3 26.4 79.0

134 23.4 12.6 37.6



#90

Hexachloroethane

Concen: 49.132 ug/l

RT: 14.335 min Scan# 2105

Delta R.T. 0.000 min

Lab File: VN079635.D




Acq: 23 Oct 2023 09:28


Tgt Ion: 117 Resp: 118915

Ion Ratio Lower Upper

117 100


201 67.5 29.3 88.0








#91
 1,2-Dichlorobenzene
 Concen: 47.549 ug/l
 RT: 14.106 min Scan# 2066
 Delta R.T. -0.006 min
 Lab File: VN079635.D
 Acq: 23 Oct 2023 09:28

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion:146 Resp: 310949
 Ion Ratio Lower Upper
 146 100
 111 42.5 22.0 66.0
 148 64.1 31.8 95.3



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 37.913 ug/l
 RT: 14.723 min Scan# 2171
 Delta R.T. 0.000 min
 Lab File: VN079635.D
 Acq: 23 Oct 2023 09:28

Tgt Ion: 75 Resp: 37792
 Ion Ratio Lower Upper
 75 100
 155 76.0 39.1 117.2
 157 95.1 50.3 150.9





#93

1,2,4-Trichlorobenzene

Concen: 50.300 ug/l

RT: 15.394 min Scan# 21

Delta R.T. -0.006 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:180 Resp: 101877

Ion Ratio Lower Upper

180 100

182 92.8 46.1 138.2

145 31.4 16.6 49.8



#94

Hexachlorobutadiene

Concen: 44.209 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.006 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28




Tgt Ion:225 Resp: 69737


Ion Ratio Lower Upper

225 100

223 63.5 31.3 93.8

227 64.1 31.9 95.9





#95

Naphthalene

Concen: 40.422 ug/l

RT: 15.641 min Scan# 21

Delta R.T. -0.006 min MSVOA_N

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:128 Resp: 250004

Ion Ratio Lower Upper

128 100

127 13.8 10.0 15.0

129 11.0 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 42.912 ug/l

RT: 15.847 min Scan# 2362

Delta R.T. 0.006 min

Lab File: VN079635.D

Acq: 23 Oct 2023 09:28

Tgt Ion:180 Resp: 103717

Ion Ratio Lower Upper

180 100

182 94.5 46.2 138.6

145 33.5 17.5 52.6

