

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
 Data File : VN079637.D
 Acq On : 23 Oct 2023 10:26
 Operator : JC\MD
 Sample : VN1023WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1023WBL02

Quant Time: Oct 23 23:45:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	264164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	515064	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	524417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	173545	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	226113	52.791	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.580%
35) Dibromofluoromethane	8.171	113	204269	55.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.380%
50) Toluene-d8	10.570	98	619111	46.227	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.460%
62) 4-Bromofluorobenzene	12.853	95	220715	51.352	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
Data File : VN079637.D
Acq On : 23 Oct 2023 10:26
Operator : JC\MD
Sample : VN1023WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1023WBL02

Quant Time: Oct 23 23:45:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

