

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
 Data File : VN079642.D
 Acq On : 23 Oct 2023 12:35
 Operator : JC\MD
 Sample : PB156545ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156545ZHE#02

Quant Time: Oct 23 23:48:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	186397	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	377597	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	386766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	89385	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	164481	54.423	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.840%		
35) Dibromofluoromethane	8.177	113	138809	51.154	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.300%		
50) Toluene-d8	10.571	98	520616	53.024	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 106.040%		
62) 4-Bromofluorobenzene	12.853	95	135685	43.062	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 86.120%		
Target Compounds						
16) Acetone	4.448	43	17723	14.785	ug/l	Qvalue 91
20) Methylene Chloride	5.283	84	16615	5.797	ug/l #	83
37) Ethyl Acetate	7.571	43	4616	1.139	ug/l #	77
43) Isopropyl Acetate	8.694	43	171329	24.132	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
Data File : VN079642.D
Acq On : 23 Oct 2023 12:35
Operator : JC\MD
Sample : PB156545ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156545ZHE#02

Quant Time: Oct 23 23:48:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

