

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
 Data File : VN079643.D
 Acq On : 23 Oct 2023 12:59
 Operator : JC\MD
 Sample : PB156545ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156545ZHE#03

Quant Time: Oct 23 23:49:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	173628	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	360900	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	371167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	93433	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	157270	55.864	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 111.720%			
35) Dibromofluoromethane	8.171	113	133697	51.550	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.100%			
50) Toluene-d8	10.571	98	491726	52.399	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.800%			
62) 4-Bromofluorobenzene	12.853	95	136387	45.287	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 90.580%			
Target Compounds						
16) Acetone	4.454	43	17696	15.848	ug/l #	76
20) Methylene Chloride	5.277	84	17337	6.567	ug/l #	80
43) Isopropyl Acetate	8.694	43	174861	25.769	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
Data File : VN079643.D
Acq On : 23 Oct 2023 12:59
Operator : JC\MD
Sample : PB156545ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156545ZHE#03

Quant Time: Oct 23 23:49:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

